

Product information

Innosept PREMIUM conc. Instrument Disinfection

1000 ml REF 4011I 5000 ml REF 4011L



Characteristics

- Liquid concentrate for the manual and regular reprocessing of medical instruments
- free of peracids non-corrosive despite high effectiveness especially mild
- hence, does not attack the surfaces of the instuments - high material compatibility
- bactericidal, tuberculocidal, fungicidal, virucidal

Product type

Instrument disinfection for the disinfection of surgical instruments, including endoscopes of all kind.

Application

Before use, first dilute the concentrate with water. Following the recommendations of the RKI (Robert Koch Institute) / BfrArM (Federal Institute for Drugs and Medical Devices), ensure that the instruments are thoroughly pre-cleaned or mechanically cleaned and, if necessary, rinsed before disinfection. Insert the cleaned instruments into the (diluted) working solution in an open condition. Completely wet all surfaces and cavities to be disinfected with the working solution. Cover tubs for instrument disinfection. After the recommended application time, remove the instruments, thoroughly rinse them with water and dry them. A final rinse with demineralised water is recommended. Replace any disinfectant solutions at least once per working day. When using the instruments, tubs, and application aids for the first time or when changing products, wash and rinse them thoroughly. Do not mix with other products (e.g. anionic surfactant, amines).

Composition

Contains < 5% cationic surfactants, 5-15% anionic surfactants, fragrances

100g contain: 4.95g 1.5-pentanedial, 12.0g (ethylenedioxy)dimethanol.

Chemical-physical Data

Density:	1.0 g/cm ³
pH value:	4.5

Application times

Bacteria as per EN 13727, EN 14561, VAH (Association for Applied Hygiene) with dirty conditions:

1.5% 15min.

1.0% 60min.

Candida albicans as per EN 13624, EN 14562, VAH with dirty conditions: 1.5% 15min.

1.0% 60min.

<u>"Limited virucidal" (enveloped viruses such as HBV,</u> HCV, HIV) as per RKI/DVV (German Association for the Control of Virus Diseases) and EN 17111 (dirty conditions):

0.5% 5min.

TB (Tuberculosis) as per EN 14348, EN 14563 (clean conditions): 3.0% 30min

1.5% 60min.

Mycobacteria (*M. avium* and *M. terrae*) as per EN 14348, EN 14563 (clean conditions): 3.0% 60min.

Non-enveloped viruses (adeno-, polio and noroviruses) as per EN 14476 and EN 17111 (clean conditions): 3.0% 60min.

Polyomaviruses (SV 40) as per RKI/DVV (dirty conditions):

1.5% 15min.

Fungi (A. brasiliensis) as per EN 13624, EN 14562 (dirty conditions):

3% 60min.

Certificates

Our products are scientifically examined and tested according to the latest methods. We will be pleased to send you copies of our expert opinions on request.

Contact

PRISMAN Pharma International AG Am Stalden 16 4622 Egerkingen Switzerland

Tel.: +41 (0)62 206 20 60 Fax: +41 (0)62 206 20 69

www.innocid.com info@prismanpharma.com

Manufacturer

PRISMAN GmbH Otto-Hahn-Ring 6-18 64653 Lorsch Germany

Safety instructions:

H 302 Harmful if swallowed. H 315 Causes skin irritation. H 318 Causes serious eye damage. H 334 May cause allergy or asthma symptoms or breathing difficulties if inhaled. H 317 May cause an allergic skin reaction. H 335-H 336 May cause respiratory irritation. May cause drowsiness or dizziness. H 400 Very toxic to aquatic life. Wear protective gloves/eye protection. IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER/doctor. If experiencing respiratory symptoms: Call a POISON CENTER/ doctor. IF SWALLOWED: Call a POISON CENTER/doctor/if you feel unwell. Avoid release to the environment. Dispose of contents/container in accordance with local/regional/national /international regulations. Contains glutaral, (ethylendioxy)- dimethanol.



